

AUTOCLAVE LOG

Use of this form is voluntary and not required by the Department of Health. The form is provided as a service to assist salons in complying with the record-keeping requirements of Chapter 64E-19, FAC.

DATE	<u>LENGTH OF RUN</u>		CHECK IF CLEANING DONE	CHECK IF SERVICING DONE	CHECK IF SPORE TESTING DONE	CHECK IF TEST RESULTS RECEIVED
	TODAY	CUMULATIVE TOTAL				

- There must be a sterilization indicator in each autoclaving to monitor the sterilization procedure. The indicator must indicate exposure to steam and 250 degrees Fahrenheit.
- Spore tests must be performed every 40 hours of autoclave operation, but not less than quarterly.
- Spore tests must be verified through an independent laboratory.
- The autoclave must be cleaned at the frequency recommended by the manufacturer.
- The autoclave must be serviced at the frequency recommended by the manufacturer, but not less than once a year.
- A copy of the manufacturer's instructions for cleaning and servicing the autoclave must be kept on file in the establishment.